

# 台架封装的激光检测头 IR

#55-298



Bench Mounted Laser Detection Head IR, #55-298

激光检测产品(上转换片)可为使用紫外、可见和红外激光用户提供更好的的产品性能和安全性。它们可以减少许多实际应用中与光束可视化、轮廓分析和对准相关的问题。每个产品系列都有三种型号。激光检测产品的层压信用卡式型号仅适用于低功率光源和反射观察。25 毫米圆盘和夹持棒式适用于需要频繁定位的部件。可拆卸的圆盘可安装位在光学元件上,以实现精确对准,而夹持棒式的型号则便于在光路中进行操作。光学可安装平台头式具有较大的活动区域可使用 1/4-20 螺纹安装,可与标准英制接杆/接杆套筒一起使用。



### 概况

类型: Detection Head

典型应用: 808nm, 820nm, 830nm, 880nm, 960 - 980nm Laser Diodes,

Nd:YAG, 1550nm Telecommunications

## 物理和机械特性

有效区直径(mm):35.00外径 (mm):70.00厚度 (mm):8.00

## 光学特性

**波**长: R

Emission Color: Green (550nm), other peaks at Red (673nm) and Blue (400nm)

Band 1: 790 - 840nm Stimulation Range: Band 2: 870 - 1070nm

Band 3: 1500 - 1590nm

Minimum Stimulation, Pulsed: 250 kW/cm<sup>2</sup> @ 1064nm, 7ns, 10Hz

### 电气性能

Persistence (Stimulation Removed): 800 µs


 $$<2\,\mu\text{W/cm}^2\@$  808nm Minimum Stimulation, Continuous:  $$<175\,\text{nW/cm}^2\@$  960nm

<100  $\mu$ W/cm $^2$  @ 1550nm Maximum Stimulation, Continuous: 100 W/cm $^2$  @ 1064nm

Maximum Stimulation, Single Pulse: 35 MW/cm<sup>2</sup> @ 1064nm, 7ns







**IR Detection Products** 





April 29, 2025

To Whom It May Concern,

This document certifies that the product stated below has been reviewed as requested by Edmund Optics:

商品编号	<i>产品名称</i>
55-298	<b>台架封装的激光</b> 检测头 IR

This item does not contain any of the substances listed on the Reach SVHC list as of 27 June 2018 in quantities exceeding 0.1% weight by weight.

Please note the following:

- Entry 11 of Annex V of the REACH regulation EC 1907/2006 as amended by regulation 987/2008, exempts "non-hazardous" glass components from registration.
- Items considered to be packing materials and which do not form part of, or are not required for the operation of, the product, are not considered to be within scope of this declaration.
- EO does not routinely analyze products for substances not purposely added or specified within the design specifications, drawings and assembly procedures.

This certification means that:

- EO's suppliers have confirmed the material composition of this product.
- EO has implemented rigorous procedures to document this compliance.
- The information provided may, or may not, be based upon actual test data, or on information from our Vendors, Raw Material Suppliers or Subcontractors.

1 Budl

Jay Budd, Director of Corporate Compliance April 29, 2025

Edmund Optics Inc. - 101 E Gloucester Pike, Barrington, NJ 08007 | 1-800-363-1992 | Compliance@edmundoptics.com





April 29, 2025

To Whom It May Concern,

This document certifies that the product stated below has been reviewed as requested by Edmund Optics:

This item is EU RoHS (2015/863/EU) compliant without the use of exemptions.

This certification means that:

- EO's suppliers have confirmed the material composition of this product.
- EO has implemented rigorous procedures to document this compliance.
- The information provided may, or may not, be based upon actual test data, or on information from our Vendors, Raw Material Suppliers or Subcontractors.

1 Book

Jay Budd, Director of Corporate Compliance April 29, 2025

Edmund Optics Inc. - 101 E Gloucester Pike, Barrington, NJ 08007 | 1-800-363-1992 | Compliance@edmundoptics.com



April 29, 2025

商品编号 产品名称

55-298 **台架封装的激光**检测头 IR

Edmund Optics certifies that all articles included in this shipment are in compliance with the terms and conditions of this order. The company also certifies that the articles included in this shipment are in accordance to all agreed upon specifications and quality assurance provisions. Please call 1-800-363-1992 should any questions arise from this shipment.

M Hay

Jeff Harvey- EVP of Operations



